















UHV-vakuumdicht geschweisst
Integral-Leckrate: kleiner 1*10^{-9} mbar*l/s
heliumdicht bei 10^{-9} mbar

ultraschallgereinigt

Kanten ISO 13715 |+0.3 -0.3 |-0.1

25 vak Rz 1

Copying of this document and giving it to others and the use of communication of the contents thereof are forbidden without express authority. Offenders are liable to the payment of damages. All rights are reserved in the event of the grant of a patent or the registration of a utility model or design.

Projection

Surfaces DIN ISO 1302

Dimensions without tolerance indication DIN ISO 2768-mK EN ISO 13920-BF

Rev. No.	Rev.-Doc.-No.	Date	Name
-			

For information

Height

Material

Scale 1:2 (1:1)

Size A0

	Date	Name
Drawn	04.12.2013	Hartung, W.
Checked		
Approved		

Description Kammer

Drawing No. STD-000-L01.501.000

Rev. -

Sheet 1 of 2

Prev. Dwg. DK 480.010 A

Appl. for